

INSTRUCTIONS

Scope Case for IPLEX G MAJ-2339

1 *Intended use*

This scope case is intended to be used in combination with with the scopes of the IPLEX G. For additional information, refer to the IPLEX G instruction manual.

2 *Package contents*

Scope case main body 1 Instructions (this leaflet)..... 1

Match all the items in the package with the components listed above. If you have any question regarding the usability of the main body or an accessory, do not use it. Contact EVIDENT.

3 *Specifications*

Item	Specifications
Dimensions	455(W) x 185(H) x 330(D) mm (excluding protrusions)
Weight	2.0 kg

4 *Handling precautions*

- Do not hold up or use the scope case while its cover is open.

5 *Storage precautions*

- Do not expose the scope case to high temperature, high humidity, direct sunlight or radioactive rays.

6 *Troubleshooting and repairs*

- If you suspect even the slightest problem, contact EVIDENT immediately.

Manufactured by

EVIDENT CORPORATION

6666 Inatomi, Tatsuno-machi, Kamiina-gun, Nagano 399-0495, Japan

Industrial Endoscope Accessory

取扱説明書

IPLEX G 用スコープケース MAJ-2339

1 本製品の使用目的

本製品は、IPLEX G のスコープと組み合わせて使用してください。
使用に際して、以下に示した項目以外については、IPLEX G の取扱説明書に従ってください。

2 梱包品の確認

スコープケース本体 1 取扱説明書(本書) 1

本体および付属品を確認してください。万一本体および付属品に疑問な点がある場合は、
使用しないで直ちにお買い上げになった販売店または当社支店、営業所にご連絡ください。

3 仕様

項目	仕様
外形寸法	455(W)×185(H)×330(D)mm (突起部を含まず)
質量	2.0kg

4 取り扱い上の注意

- ふたが閉まっていない状態でスコープケースを持ち上げたり使用したりしないでください。

5 保管上の注意

- 高温・高湿および直射日光や放射線の当たる場所には置かないでください。

6 故障、修理について

- 少しでも不具合を感じた場合には、直ちにお買い上げになった販売店または当社支店、営業所にお問い合わせください。

株式会社エビデント



EVIDENT Customer Information Center

お客様相談センター

0120-58-0414 受付時間 平日 9:00～17:00

※携帯・PHSからもご利用になれます。

生物・工業用顕微鏡 E-mail: ot-cic-microscope@evidentscientific.com

工業用内視鏡 E-mail: ot-cic-inspro@evidentscientific.com

工業用内視鏡アクセサリ

使用说明书

重要: 使用本机之前请务必阅读此使用说明书。

IPLEX G 专用内窥镜插入管箱 MAJ-2339

1 用途

本产品设计与 IPLEX G 一起使用。
有关更多信息, 请参阅 IPLEX G 使用说明书。

2 包装内容

内窥镜插入管箱主体1 说明书 (本页)1

请将包装内的物品与上面所列的部件进行比对。如对主体或配件有任何疑问, 请勿继续使用, 应与仪景通联系。

3 规格

项目	规格
尺寸	455(W) x 185(H) x 330(D) mm (不包括突出部)
重量	2.0 kg

4 操作注意事项

- 在没有关闭盖子的状态下, 切勿拿起或使用内窥镜插入管箱。

5 存放注意事项

- 请勿将内窥镜插入管箱置于高温的、高湿度的、阳光直射的或放射线照射的地方。

6 故障排除与维修

- 不管任何现象, 如果发现了异常情况, 务请立刻与仪景通联系。

销售商名称 - 地址

生产日期: 请见捆包箱

仪景通光学科技 (上海) 有限公司

总部: 中国 (上海) 自由贸易试验区日樱北路 199-9 号 102 及 302 部位
北京分公司: 北京市朝阳区酒仙桥路 10 号 1 幢一层 102 室
售后服务热线: 400-969-0456

2022年9月发行

工业内窥镜附件